

# CHAPTER 1

## PRE-INSTALLATION

### CONTENTS:

1.1.0	INTRODUCTION .....	1-2
1.1.1	Terminology .....	1-2
1.1.2	Generator Description .....	1-2
1.1.3	Features .....	1-3
1.1.4	Radiographic Performance .....	1-3
1.1.5	Environmental Specifications .....	1-4
1.1.6	Applicable Standards .....	1-5
1.1.7	Electromagnetic Compatibility (EMC).....	1-8
1.1.8	RoHS Compliance.....	1-9
1.1.9	Product Marking .....	1-9
1.2.0	SAFETY .....	1-10
1.2.1	Safety and Warning Symbols.....	1-10
1.2.2	Safety Notices and Warnings.....	1-10
1.2.3	Safety Warning Labels .....	1-13
1.3.0	PREPARING FOR INSTALLATION.....	1-17
1.3.1	Generator Heat Output.....	1-17
1.3.2	Generator Power Requirements .....	1-17
1.3.3	Generator Ground Requirements.....	1-21
1.3.4	X-Ray Stator Drive Cable Requirements .....	1-21
1.3.5	Locating and Mounting the Generator.....	1-22
1.3.6	Dimensions, Cable Entrance and Seismic Center Location.....	1-23
1.3.7	Tools and Test Equipment Required.....	1-24
1.3.8	Pre-Installation Checklist.....	1-25
1.4.0	GENERATOR LAYOUT AND MAJOR COMPONENTS .....	1-25
1.5.0	COMPATIBILITY LISTING.....	1-28
1.6.0	RUNTIME LICENSE AGREEMENT (TOUCHSCREEN CONSOLE).....	1-29
1.7.0	CUSTOMER SUPPORT .....	1-30
1.8.0	COMPATIBILITY STATEMENT .....	1-30
1.9.0	OPEN-SOURCE LICENSE AGREEMENT (TOUCHSCREEN CONSOLE).....	1-31

Use and disclosure is subject to the restrictions on the title page of this CPI document.

**1.1.0 INTRODUCTION**

This chapter summarizes the main features of the CMP 200<sup>®</sup> and CMP 200<sup>®</sup> DR X-ray generators (performance, regulatory and compatibility). Safety information is provided, along with environmental, room, and installation requirements. This chapter concludes with a pre-installation checklist and a diagram showing the major component layout.

The information in this chapter is provided in order for the installer to be able to plan the site layout prior to installation of the generator.

**1.1.1 Terminology**

Direct Radiography	Radiography in which the permanent recording is effected at an image reception area (i.e. film).
Indirect Radiography	Radiography in which the permanent recording is effected after transfer of the information obtained at an image reception area (i.e. digital imaging system).
Direct Radioscopy	Radioscopy in which the visible images are presented at the image reception area, or close to it, in the radiation beam (i.e. fluoroscope or image intensifier with a non-digital imaging system).
Indirect Radioscopy	Radioscopy in which the images are presented at a location outside the radiation beam after transfer of the information (i.e. digital imaging system with a flat-panel detector or with an image intensifier and a CCD camera).
Reference Air Kerma Rate	Air Kerma rate free in air in the primary X-ray beam measured under specific conditions and expressed at the patient entrance reference point.
Patient Entrance Reference Point	Point intended to represent the intersection of the X-ray beam axis with the entrance surface of the patient.
Isocentre	In radiological equipment with several modes of movement of the reference axis around a common centre, the centre of the smallest sphere through which the radiation beam axis passes.
Entrance Field Size	Dimensions of the field in the entrance plane of an X-ray image receptor that can be used for the transmission of an X-ray pattern under specific conditions.

**1.1.2 Generator Description**

The CMP 200<sup>®</sup> 100 kHz high frequency X-ray generator is a component for use in film-based stationary radiographic X-ray systems. The CMP 200<sup>®</sup> DR X-ray generator adds a digital interface for digital radiography (DR) equipment. The CMP 200<sup>®</sup> X-ray generator consists of a main power cabinet and an optional membrane control console. The CMP 200<sup>®</sup> DR X-ray generator consists of a main power cabinet and an optional membrane, touchscreen or mini-console (used with digital interface). The main power cabinet contains the High Voltage Module and control circuits, the filament drivers, a low speed starter (optional dual-speed starter on some models of CMP 200<sup>®</sup> DR), and interface connections to the room equipment.

The control console allows the operator to select the technique factors, image receptors, etc., and to initiate an X-ray exposure.

Use and disclosure is subject to the restrictions on the title page of this CPI document.

### 1.1.3 Features

The following are the main features of and the options available for the generator:

- Integral low speed starter, compatible with X-ray tubes with type "R" stator. Optional compatibility with GE 23/23  $\Omega$  equal impedance "E" stator.
- Optional dual-speed starter on some models of CMP 200<sup>®</sup> DR (not available on 208 / 230 VAC, 1-phase & 3-phase units), compatible with tube types listed in chapter 2.
- Capable of interfacing with various DR imaging systems (CMP 200<sup>®</sup> DR only).
- 24 VDC, 110, or 220 VAC power source for Buckys (fused at 0.8 amps).
- 24 VAC 150 watts power source for collimator lamp.
- 24 VDC 45 watts power source for system locks.
- Optional AEC.
- Optional DAP (Dose-Area Product).
- Tomography

### 1.1.4 Radiographic Performance

kVp range:	40 to 125 kV or 40 to 150 kV, depending on model.
kVp steps:	variable in 1 kV steps.
kVp accuracy:	$\pm 5\% + 1 \text{ kV}$ .
Risetime (10-90%):	<1.5 ms.
Time range:	1.0 to 6300 milliseconds.
mAs range	0.1 to 500 mAs (32/40 kW) 0.1 to 630 mAs (50 kW) 0.1 to 800 mAs (65 kW). 0.1 to 1000 mAs (80 kW)
mAs accuracy:	$\pm (10\% + 0.2) \text{ mAs}$ .
mA range	10 to 400 mA (32 kW) 10 to 500 mA (40 kW) 10 to 630 mA (50 kW) 10 to 800 mA (65 kW) 10 to 1000 mA (80 kW)
Coefficient of linearity:	0.05 (station to station) mAs.
Coefficient of reproducibility:	kV, mAs $\leq 0.05$ .
Duty Cycle	Not to exceed 5 consecutive boosts, followed by a minimum 10 second wait period.

Use and disclosure is subject to the restrictions on the title page of this CPI document.

**1.1.5 Environmental Specifications****OPERATING**

Ambient temperature range	10 to 40 °C (50 to 104 °F).
Relative humidity	20 to 80%, non-condensing.
Atmospheric pressure range	1060 to 700 hPa (-400 to +3000 meters, 795 to 525 mm Hg); Reference: 1013 hPa nominal at sea level.

**TRANSPORT AND STORAGE**

Ambient temperature range	-20 to 70 °C (-13 to 158 °F)..
Relative humidity	5 to 95%, non-condensing.
Atmospheric pressure range	1060 to 700 hPa (-400 to +3000 meters, 795 to 525 mm Hg); Reference: 1013 hPa nominal at sea level.

- Electrolytic capacitors contained within the equipment require less than +40°C for long-term storage life.
- The membrane control console is limited to a minimum temperature of -20°C, with a maximum duration of 48 hours at that temperature. Transport and storage is limited to a maximum duration of 120 hours between 50 and 70 °C, with an absolute humidity not to exceed the humidity of 85% RH at 50 °C.
- Touchscreen console temperatures below -20°C and above +50°C are limited to 10 days maximum duration, with a humidity not exceeding 50 % RH.

Use and disclosure is subject to the restrictions on the title page of this CPI document.

### 1.1.6 Applicable Standards

The CMP 200® / CMP 200® DR series of X-ray generators comply with the regulatory requirements and design standards in this section as follows:

- VZW2555 series: Only the standards marked with an asterisk \* under **SAFETY**.
- VZW2556 series: All standards in this section.

#### A) SAFETY

- \* FDA Center for Devices & Radiological Health (CDRH) - 21 CFR subchapter J (USA), Part 1010 and 1020, Class I.
- \* Radiation Emitting Devices Act - C34 (Canada).
- Medical Device Regulations (Canada), Class 3.
- EC Directive 93/42/EEC (amended by 2007/47/EC) concerning Medical Devices (European Community), Class IIb.
- EU Commission Regulation 207/2012 on electronic labeling of medical devices.
- \* EN 60601-1/IEC 60601-1, EN 60601-2-7/IEC 60601-2-7, CSA 60601.1, UL 60601.1.
- CAN/CSA-C22.2 No. 60601-1-08, IEC 60601-1:2005/EN 60601-1:2006, CAN/CSA 60601-2-54-11, IEC 60601-2-54:2009/EN 60601-2-54:2009, ANSI/AAMI ES60601-1:2005.
  - -Type of protection against electric shock: Class I equipment. **Warning: To avoid the risk of electric shock, this equipment must only be connected to a supply mains with protective earth.**
  - -Overvoltage category classification: II.
  - -Pollution degree classification: 2.
  - -Degree of protection against electric shock: Not classified.
  - -Degree of protection against harmful ingress of water: Ordinary equipment.
  - -Mode of operation: Continuous operation with intermittent loading (standby - exposure).
  - -Equipment not suitable for use in presence of a flammable anesthetic mixture with air or with oxygen or nitrous oxide.
- EN 62304/IEC 62304 - Software life-cycle processes.
- EN ISO 14971 - Risk Management.
- EN 1041 - Information supplied by the manufacturer of medical devices.
- EN 60601-1-6/IEC 60601-1-6. EN 62366/IEC 62366 – Usability and application of usability engineering to medical devices.

Use and disclosure is subject to the restrictions on the title page of this CPI document.

## 1.1.6 Applicable Standards (Cont)

- EN 60601.1.2/IEC 60601.1.2, applying the following standards:

## Electromagnetic Immunity

IEC61000-4-2	Electrostatic Discharge
IEC61000-4-3	Radiated RF Field
IEC61000-4-4	Electrical Fast Transient
IEC61000-4-5	Surge
IEC61000-4-6	Conducted RF Immunity
IEC61000-4-8	Magnetic Field Immunity

## Electromagnetic Emission:

EN 55011 (CISPR Publication II Emissions Standards Group 1, Class A)

NOTE: All referenced standards are considered to be at the latest adopted revision.

## B) EMC (EN 60601-1-2:2001/IEC 60601-1-2:2001)

Guidance and manufacturer's declaration – electromagnetic emissions		
The VZW2556 series of X-ray generators is intended for use in the electromagnetic environment specified below. The customer or the user of the VZW2556 series should assure that it is used in such an environment.		
Emissions test	Compliance	Electromagnetic environment - guidance
RF emissions CISPR 11	Group 1	The VZW2556 series of X-ray generators use RF energy only for their internal functions. Therefore, the RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.
RF emissions CISPR 11	Class A (The VZW2556 series of X-ray generators in combination with shielded location)	The VZW2556 series of X-ray generators must be used only in a shielded location with a minimum RF shielding effectiveness and, for each cable that exits the shielded location, a minimum RF filter attenuation of 40dB from 30 MHz to 230 MHz and 47dB from 230 MHz to 1 GHz. (The minimum at 30 MHz is 40dB and the minimum at 230 MHz is 47dB).
Harmonic emissions IEC 61000-3-2	Not Applicable	The VZW2556 series is suitable for use in all establishments other than domestic and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.
Voltage fluctuations/ flicker emissions IEC 61000-3-3	Not Applicable	
NOTE It is essential that the actual shielding effectiveness and filter attenuation of the shielded location be verified to assure that they meet the minimum specifications.		


Use and disclosure is subject to the restrictions on the title page of this CPI document.

## 1.1.6 Applicable Standards (Cont)

Guidance and manufacturer's declaration – electromagnetic immunity			
The VZW2556 series of X-ray generators is intended for use in the electromagnetic environment specified below. The customer or the user of the VZW2556 series should assure that it is used in such an environment.			
Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment – guidance
Electrostatic discharge (ESD) IEC 61000-4-2	$\pm 6$ kV contact $\pm 8$ kV air	$\pm 6$ kV contact $\pm 8$ kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30%.
Electrical fast transient/burst IEC 61000-4-4	$\pm 2$ kV for power supply lines $\pm 1$ kV for input/output lines	$\pm 2$ kV for power supply lines $\pm 1$ kV for input/output lines	Mains power quality should be that of a typical commercial or hospital environment.
Surge IEC 61000-4-5	$\pm 1$ kV line to line $\pm 2$ kV line to ground	$\pm 1$ kV line to line $\pm 2$ kV line to ground	Mains power quality should be that of a typical commercial or hospital environment.
Voltage dips, short interruption, and voltage variations on power supply input lines IEC 61000-4-11	$< 5\% U_T$ ( $> 95\%$ dip in $U_T$ ) for 0.5 cycle  $40\% U_T$ (60 % dip in $U_T$ ) for 5 cycles  $70\% U_T$ (30 % dip in $U_T$ )  $< 5\% U_T$ ( $> 95\%$ dip in $U_T$ ) for 5 s	$< 5\% U_T$ ( $> 95\%$ dip in $U_T$ ) for 0.5 cycle  $40\% U_T$ (60 % dip in $U_T$ ) for 5 cycles  $70\% U_T$ (30 % dip in $U_T$ )  $< 5\% U_T$ ( $> 95\%$ dip in $U_T$ ) for 5 s	Mains power quality should be that of a typical commercial or hospital environment. If the user of the VZW2556 series X-ray generator requires continued operation during power mains interruptions, it is recommended that the X-ray generator be powered from an uninterruptible power supply or battery.
Power frequency (50/60 Hz) IEC 61000-4-8	3 A/m	3 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment
NOTE: $U_T$ is the A.C. mains voltage prior to application of the test level.			

Use and disclosure is subject to the restrictions on the title page of this CPI document.

## 1.1.6 Applicable Standards (Cont)

Guidance and manufacturer's declaration – electromagnetic immunity			
The VZW2556 series of X-ray generators is intended for use in the electromagnetic environment specified below. The customer or the user of the VZW2556 series should assure that it is used in such an environment.			
Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment - guidance
Conducted RF IEC 61000-4-6	3 V <sub>rms</sub> 150 kHz to 80MHz	3 V <sub>rms</sub> 150 kHz to 80MHz	The VZW2556 series of X-ray generators must be used only in a shielded location with a minimum RF shielding effectiveness and, for each cable that enters the shielded location, a minimum RF filter attenuation of 40dB from 30 MHz to 230 MHz and 47dB from 230 MHz to 1 GHz. (The minimum at 30 MHz is 40dB and the minimum at 230 MHz is 47dB.)
Radiated RF IEC 61000-4-3	3 V/m 80MHz to 2.5 GHz	3 V/m 80MHz to 2.5 GHz	Field strengths outside the shielded location from fixed RF transmitters, as determined by an electromagnetic site survey, should be less than 3 V/m. <sup>a</sup>  Interference may occur in the vicinity of equipment marked with the following symbol:  
NOTE 1 These guidelines may not apply all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.			
NOTE 2 It is essential that the actual shielding effectiveness and filter attenuation of the shielded location be verified to assure that they meet the minimum specification.			
<sup>a</sup> Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the VZW2556 series of X-ray generators is used exceeds the applicable RF compliance level above, the X-ray generator should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as re-orienting or relocating the X-ray generator.			

## 1.1.7 Electromagnetic Compatibility (EMC)

In accordance with the intended use, some models of this series of X-ray generators comply with the European Council Directive concerning Medical Devices. The CE marking affixed to compliant products signifies this. One of the harmonized standards of this Directive defines the permitted levels of electromagnetic emission from this equipment and its required immunity from the electromagnetic emissions of other devices.

It is not possible, however, to exclude with absolute certainty the possibility that other high frequency electronic equipment, which is fully compliant to the EMC regulations, will not adversely affect the operation of this generator. If the other equipment has a comparatively high level of transmission power and is in close proximity to the generator, these EMC concerns (the risk of interference) may be more pronounced. It is therefore recommended that the operation of equipment of this type such as mobile telephones, cordless microphones and other similar mobile radio equipment be restricted from the vicinity of this X-ray generator.

Use and disclosure is subject to the restrictions on the title page of this CPI document.



### 1.1.8 RoHS Compliance

ACPEIP-----China RoHS

The following are not specific design requirements at this time but are objectives for consideration over the life of the product.

DIRECTIVE 2011/65/EU, RoHS-----Restriction of Hazardous Substances

CMP 200® / CMP 200® DR  
Table of hazardous substances' name and concentration

#### 产品中有毒有害物质或元素的名称及含量

Component Name 部件名称	Hazardous substances' name 有毒有害物质或元素					
	Lead	Mercury	Cadmium	Hexavalent Chromium	Poly-Brominated Biphenyls	Poly-Brominated Diphenyl Ether
	(Pb)	(Hg)	(Cd)	(Cr(VI))	(PBB)	(PBDE)
	铅	汞	镉	六价铬	多溴联苯	多溴二苯醚
Generator	X	O	O	X	O	O
Console	X	O	O	X	O	O

O: Indicates that this toxic or hazardous substance contained in all of the homogeneous materials for this part is below the limit requirement in SJ/T11363-2006.  
X: Indicates that this toxic or hazardous substance contained in at least one of the homogeneous materials used for this part is above the limit requirement in SJ/T11363-2006.

- Data listed in the table represents best information available at the time of publication
- Applications of hazardous substances in this medical device are required to achieve its intended clinical uses, and/or to provide better protection to human beings and/or to environment, due to lack of reasonably (economically or technically) available substitutes.

O: 表示该有毒有害物质在该部件所有均质材料中的含量均在SJ/T11363-2006 标准规定的限量要求以下  
X: 表示该有毒有害物质至少在该部件的某一均质材料中的含量超出SJ/T11363-2006 标准规定的限量要求

- 此表所列数据为发布时所能获得的最佳信息
- 由于缺少经济上或技术上合理可行的替代物质或方案, 此医疗设备运用以上一些有毒有害物质来实现设备的预期临床功能, 或给人员或环境提供更好的保护效果。

### 1.1.9 Product Marking



The CE Mark is a declaration by the manufacturer that the product complies with the requirements of the applicable European Union (EU) medical device directive and that the product has been subject to conformity assessment procedures as provided in that directive.






A CSA mark with the indicators "C" and "US" means that product is certified for both the U.S. and Canadian markets, to the applicable U.S. and Canadian standards.

Use and disclosure is subject to the restrictions on the title page of this CPI document.

## 1.2.0 SAFETY

## 1.2.1 Safety and Warning Symbols

The following advisory symbols are used on the safety warning labels, and/or on circuit boards, and/or on the operator console.

	High voltage symbol used to indicate the presence of high voltage.
	Warning symbol used to indicate a potential hazard to operators, service personnel or to the equipment. It indicates a requirement to refer to the accompanying documentation for details.
	Radiation exposure symbol used on operator console. Lights to indicate that an exposure is in progress. This is accompanied by an audible tone from the console.
<b>WARNING</b> <i>THIS X-RAY UNIT MAY BE DANGEROUS TO PATIENT AND OPERATOR UNLESS SAFE EXPOSURE FACTORS, OPERATING INSTRUCTIONS AND MAINTENANCE SCHEDULES ARE OBSERVED.</i>	Radiation warning label on operator console.  Never allow unqualified personnel to operate the X-ray generator.

## 1.2.2 Safety Notices and Warnings

**WARNING:**    *THIS X-RAY UNIT MAY BE DANGEROUS TO PATIENT AND OPERATOR UNLESS SAFE EXPOSURE FACTORS, OPERATING INSTRUCTIONS AND MAINTENANCE SCHEDULES ARE OBSERVED.*

**WARNING:**    *PROPER USE AND SAFE OPERATING PRACTICES WITH RESPECT TO X-RAY GENERATORS ARE THE RESPONSIBILITY OF USERS OF SUCH GENERATORS. CPI CANADA INC. ("THE MANUFACTURER") PROVIDES INFORMATION ON ITS PRODUCTS AND ASSOCIATED HAZARDS, BUT ASSUMES NO RESPONSIBILITIES FOR AFTER-SALE OPERATING AND SAFETY PRACTICES.*

*THE MANUFACTURER ACCEPTS NO RESPONSIBILITY FOR ANY GENERATOR NOT MAINTAINED OR SERVICED ACCORDING TO THIS SERVICE AND INSTALLATION MANUAL, OR FOR ANY GENERATOR THAT HAS BEEN MODIFIED IN ANY WAY.*

*THE MANUFACTURER ALSO ASSUMES NO RESPONSIBILITY FOR X-RAY RADIATION OVEREXPOSURE OF PATIENTS OR PERSONNEL RESULTING FROM POOR OPERATING TECHNIQUES OR PROCEDURES.*

Use and disclosure is subject to the restrictions on the title page of this CPI document.

### 1.2.2 Safety Notices and Warnings (Cont)

X-ray radiation exposure may be damaging to health, with some effects being cumulative and extending over periods of many months or even years. **Operators and service personnel should avoid any exposure to the primary beam** and take protective measures to safeguard against scatter radiation. Scatter radiation is caused by any object in the path of the primary beam and may be of equal or less intensity than the primary beam that exposes the film.

No practical design can incorporate complete protection for operators or service personnel who do not take adequate safety precautions. **Only authorized and properly trained service and operating personnel should be allowed to work with this X-ray generator equipment.** The appropriate personnel must be made aware of the inherent dangers associated with the servicing of high voltage equipment and the danger of excessive exposure to X-ray radiation during system operation.



#### **DO NOT CONNECT UNAPPROVED EQUIPMENT TO THE REAR OF THE CONSOLE.**

For the membrane console, J3 is for connection of an external hand switch, J4 is a serial port for use by an external computer, and J8 is for the interconnect cable to the main cabinet.

For the touchscreen console, GEN on the rear of the touchscreen is for the interconnect cable to the generator, HS is for connection of an external hand switch, COM 1 & COM 2 are serial ports for use by external devices, LO (3.5 mm stereo jack) is for customer supplied speakers (minimum 8 ohms, do not use externally amplified speakers), ETH is a standard 10/100 Ethernet connection, USBA and USBB are USB ports for connection of external devices, CF is for the compact flash memory card which holds the touchscreen software, and SW1 is the console upgrade button.

#### **INCORRECT CONNECTIONS OR USE OF UNAPPROVED EQUIPMENT MAY RESULT IN INJURY OR EQUIPMENT DAMAGE.**

**CAUTION: DO NOT EXCEED THE TUBE MAXIMUM OPERATING LIMITS. INTENDED LIFE AND RELIABILITY WILL NOT BE OBTAINED UNLESS GENERATORS ARE OPERATED WITHIN PUBLISHED SPECIFICATIONS.**

**NOTE: THE INSTALLER MUST PROVIDE A VISUAL INDICATION OF THE ON/OFF STATE OF EACH EXTERNAL DEVICE THAT CAN PREVENT THE GENERATOR FROM EMITTING RADIATION, OR THAT CAN STOP THE GENERATOR FROM EMITTING RADIATION, OR BOTH.**

**WARNING: DO NOT MODIFY THIS EQUIPMENT WITHOUT AUTHORIZATION OF THE MANUFACTURER.**

**NOTE: ALL ELECTRICAL WORK PERFORMED DURING INSTALLATION AND SERVICE OF THIS X-RAY GENERATOR MUST BE PERFORMED IN ACCORDANCE WITH CSA STANDARD Z462 OR EQUIVALENT.**

Use and disclosure is subject to the restrictions on the title page of this CPI document.

## 1.2.2 Safety Notices and Warnings (Cont)

**WARNING:**     **HAZARDOUS VOLTAGES EXIST INSIDE THE GENERATOR WHENEVER THE MAIN POWER DISCONNECT IS SWITCHED ON. THESE AREAS INCLUDE, BUT ARE NOT LIMITED TO, THE MAIN FUSE HOLDER AND ASSOCIATED CIRCUITS ON THE H.V. AUXILIARY BOARD, THE AUXILIARY TRANSFORMER, AND THE MAIN POWER CONTACTOR.**  
**LED DS1 ON THE H.V. AUXILIARY BOARD INDICATES THE PRESENCE OF THE +24 VDC SUPPLY.**  
**THE CONSOLE ON/OFF SWITCH DOES NOT DISCONNECT THE MAIN POWER FROM THE ABOVE AREAS INSIDE THE GENERATOR.**  
**THE DC BUS CAPACITORS, LOCATED IN THE MAIN CABINET PRESENT A SAFETY HAZARD FOR AT LEAST 5 MINUTES AFTER THE POWER HAS BEEN REMOVED FROM THE UNIT. CHECK THAT THESE CAPACITORS ARE DISCHARGED BEFORE SERVICING THE GENERATOR.**  
**AN LED CONNECTED ACROSS THE DC BUS INDICATES THE PRESENCE OF HIGH VOLTAGE. THIS LED IS MOUNTED ON THE EMC CAPACITOR BOARD (ON SOME MODELS, THE EMC CAPACITOR BOARD DOES NOT CONTAIN ANY COMPONENTS OTHER THAN THE LED AND THE SERIES RESISTORS).**  
**DO NOT RELY SOLELY ON BLEEDER CIRCUITS AND HIGH-VOLTAGE ON INDICATORS IN THE GENERATOR TO PROTECT YOU. DUE TO THE POSSIBILITY OF COMPONENT FAILURE, IT MUST NEVER BE ASSUMED THAT AN UNLIT LED ENSURES THAT NO HIGH VOLTAGE IS PRESENT. USING A VOLTMETER, CONFIRM THAT NO HIGH VOLTAGE IS PRESENT BEFORE ATTEMPTING ANY SERVICE.**

*The following notes apply to the touchscreen console only.*

**WARNING:**     **THE TOUCHSCREEN CONSOLE HAS NO USER SERVICEABLE PARTS. DO NOT ATTEMPT TO OPEN THE TOUCHSCREEN CONSOLE.**

**NOTE:**     **WHEN ATTACHING THE BASE TO THE TOUCHSCREEN CONSOLE, ENSURE THAT THE TOUCHSCREEN IS RESTING ON A FLAT, CLEAN SURFACE WITH A PIECE OF NON-ABRASIVE MATERIAL BETWEEN THE TOUCHSCREEN AND THE SURFACE.**

**NOTE:**     **TO AVOID ACCIDENTAL CONTACT WITH ENERGIZED CIRCUITRY INSIDE THE TOUCHSCREEN CONSOLE, THE MAXIMUM BACK-PLATE SCREW LENGTH MUST BE LIMITED TO 25 mm (1 INCH).**

**NOTE:**     **FOR WALL MOUNTED TOUCHSCREEN CONSOLES, ENSURE THAT THE BASE IS SECURED PROPERLY TO A WALL STUD.**

**WARNING:**     **PLEASE ENSURE THERE IS SUFFICIENT AREA AROUND THE VENTING SLOTS OF THE TOUCHSCREEN CONSOLE TO ALLOW PROPER COOLING OF THE INTERNAL COMPONENTS.**

Use and disclosure is subject to the restrictions on the title page of this CPI document.

### 1.2.2 Safety Notices and Warnings (Cont)

**NOTE:** THE SET SCREW COLLAR MUST BE POSITIONED ON THE UPPER HALF OF THE TILT ARM TO PREVENT PERSONAL INJURY SHOULD THE TOUCHSCREEN SLIP WHILE ADJUSTING THE VIEWING HEIGHT.  
USE THE PROVIDED ALLEN KEY TO ENSURE THAT THE SET SCREW COLLAR IS LOCKED INTO POSITION SUCH THAT THERE IS NO LESS THAN 25 mm (1 INCH) OF CLEARANCE BETWEEN THE BOTTOM EDGE OF THE TOUCHSCREEN CONSOLE AND THE TOUCHSCREEN BASE PLATE WHEN THE TOUCHSCREEN IS ADJUSTED TO ITS MINIMUM HEIGHT.  
PLEASE BE SURE TO SUPPORT THE TOUCHSCREEN CONSOLE WHEN ADJUSTING ITS VIEWING POSITION.

### 1.2.3 Safety Warning Labels

This subsection defines the safety labels used inside and outside the generator cabinet.

**NOTE:** THESE LABELS AND WARNINGS ARE PROVIDED TO ALERT SERVICE PERSONNEL THAT SERIOUS INJURY WILL RESULT IF THE HAZARD IDENTIFIED IS IGNORED.

**WARNING:** SWITCH OFF THE MAIN POWER DISCONNECT AND ALLOW SUFFICIENT TIME FOR ALL CAPACITORS TO DISCHARGE BEFORE REMOVING ANY COVERS.

**WARNING:** IF ANY COVERS MUST BE REMOVED FOR SERVICE, TAKE ALL REQUIRED PRECAUTIONS WITH RESPECT TO THE HAZARD(S) AND IMMEDIATELY REPLACE THE COVERS WHEN THE NEED FOR REMOVAL IS COMPLETED.



**REPLACE ALL FUSES IN THIS GENERATOR WITH THE SAME TYPE AND RATING.**


Refer to the spares list (chapter 8) for fuse replacement information.




Refer to chapter 6 for the battery replacement procedure.

This information is provided to help you establish safe operating conditions for both you and your X-ray generator. Do not operate this X-ray generator except in accordance with these instructions, and any additional information provided by the X-ray generator manufacturer and / or competent safety authorities.

Use and disclosure is subject to the restrictions on the title page of this CPI document.

**1.2.3 Safety Warning Labels (Cont)****Weight Label**

This label is attached to the main generator cabinet and to the High Voltage Module, and states the approximate weight of the main cabinet with the High Voltage Module, and the weight of the High Voltage Module separately. Do not attempt to lift these items without proper assistance.


**Caution HV/High Energy Warning Label**

This label is attached to the generator cabinet and on the inside of the back cover above the High Voltage Module. The DC bus capacitors (approximately 300 to 670 VDC, depending on model) will remain charged for up to 5 minutes after the AC mains is disconnected or the console is switched off.

Use and disclosure is subject to the restrictions on the title page of this CPI document.

### 1.2.3 Safety Warning Labels (Cont)


#### Caution HV Behind Cover Label



This label is attached to the outside of the generator cabinet, the cover over the inverter assembly, the cover over the DSS board and the fan cover. Mains voltage is present inside the cabinet whenever the main disconnect is switched on. Additionally, the DC bus capacitors will remain charged for up to 5 minutes after the AC mains is disconnected or the console is switched off.

**WARNING:    WAIT A MINIMUM OF 5 MINUTES AFTER THE INPUT MAINS POWER HAS BEEN REMOVED BEFORE REMOVING ANY COVERS. ONCE THE COVER(S) ARE REMOVED. CHECK THAT THE VOLTAGE ACROSS THE DC BUS CAPACITORS IS NEAR ZERO BEFORE SERVICING. IF THIS VOLTAGE EXCEEDS 48 VDC, THE CAPACITORS MUST BE MANUALLY DISCHARGED BY QUALIFIED SERVICE PERSONNEL.**

#### High Voltage Module - Transformer Terminals Notice



This notice is printed on the High Voltage Module lid and cautions against over-tightening the nuts on the transformer feedthrough terminals (for the primary of the HV transformers).

Use and disclosure is subject to the restrictions on the title page of this CPI document.

## 1.2.3 Safety Warning Labels (Cont)

Danger High Voltage Notice**DANGER  
HIGH TENSION**

This notice is printed on the High Voltage Module lid. High voltage may be present at the primary terminals on the tank lid board, at the output high voltage connectors, and at the mA/mAs measuring jacks if the shorting link is opened for mA/mAs measurements.

Auxiliary Transformer Labels

These labels are fixed on the outer face of the auxiliary transformer and indicate the presence of high voltage taps on the primary of 120, 200, 240, 400 and 480 VAC. Ensure the main power disconnect is switched off and appropriate documentation is consulted before attempting to service this component.



**HIGH VOLTAGE HAZARD:** Be certain that you are aware of all potential high voltage locations and hazards as detailed in this section before removing any covers, or attempting any service on this X-ray generator.



**HIGH VOLTAGE HAZARD:** Approximately 400 VAC is present on the membrane console board in the area of T1, C36, and J5. This is a high voltage source for the fluorescent backlight on the LCD display.

**HIGH VOLTAGE HAZARD:** AC mains voltage and / or DC bus voltage (approximately 325 to 670 VDC, depending on model) is present on the H.V. auxiliary board whenever the AC mains is energized. Ensure that the AC mains is switched off and locked out before servicing this board. See the note below regarding the DC bus voltage.

**HIGH VOLTAGE HAZARD:** High voltage is present on all components connected to the AC mains (line fuses, auxiliary transformer, H.V. auxiliary board, main power contactor, etc) whenever the AC mains is switched on. Additionally, DC bus voltage is present on certain components (mains rectifier assembly, DC bus capacitors, inverter assembly, High Voltage Module, H.V. auxiliary board, etc) whenever the generator is switched on, and will remain on for up to 5 minutes after the console is switched off or the AC mains is switched off or disconnected.

**HIGH VOLTAGE HAZARD:** Approximately 600 VDC is present on the dual-speed starter board whenever the generator is switched on. This voltage is sourced from the DC bus capacitors in the generator, and therefore the high voltage hazard will remain for up to 5 minutes after the generator has been switched off.

Use and disclosure is subject to the restrictions on the title page of this CPI document.



### 1.2.3 Safety Warning Labels (Cont)



High voltage (approximately 325 to 670 VDC, depending on model) is present on the inverter assembly and associated components whenever the AC mains is energized and the console is switched on, and for up to 5 minutes after the console is switched off or the AC mains is disconnected. **THIS COMBINATION OF HIGH VOLTAGE AND HIGH CURRENT IS POTENTIALLY LETHAL. USE EXTREME CAUTION WHEN SERVICING THIS UNIT.**

### 1.3.0 PREPARING FOR INSTALLATION

#### 1.3.1 Generator Heat Output

The maximum heat output of the main generator cabinet is less than 1000 BTU / hour in normal clinical use, with a maximum of 70 BTU / hour heat output for the console. The console is convection-cooled, and the main cabinet is fan cooled. The console and main cabinet should never be covered when the generator is switched on, as any covering may interfere with the cooling.

#### 1.3.2 Generator Power Requirements

The tables in this section show mains power requirements for various configurations of CMP 200<sup>®</sup> and CMP 200<sup>®</sup> DR X-ray generators. The installer must ensure that the generator is connected to the proper mains voltage as per the nameplate on the generator.

The table below depicts the power requirements for the 32 kW CMP 200<sup>®</sup> / CMP 200<sup>®</sup> DR X-ray generators.

Line Voltage	208 VAC - 5% to 230 VAC + 10%, 1 phase.
	208 VAC - 5% to 230 VAC + 10%, 3 phase.
	400 VAC $\pm$ 10%, 3 phase.
	480 VAC $\pm$ 10%, 3 phase.
Line Frequency	50/60 Hz.
Momentary Current	220 Amps at 208 VAC (1 phase).
	122 Amps/phase at 208 VAC (3 phase).
	200 Amps at 230 VAC (1 phase).
	110 Amps/phase at 230 VAC (3 phase).
	65 Amps/phase at 400 VAC.
Nominal Current *	55 Amps/phase at 480 VAC.
	$\leq$ 5 Amps.
Momentary Power Consumption	45 kVA.

Use and disclosure is subject to the restrictions on the title page of this CPI document.

**1.3.2 Generator Power Requirements (Cont)**

The table below depicts the power requirements for the 40 kW CMP 200<sup>®</sup> / CMP 200<sup>®</sup> DR X-ray generators.

Line Voltage	208 VAC - 5% to 230 VAC + 10%, 1 phase. 208 VAC - 5% to 230 VAC + 10%, 3 phase. 400 VAC $\pm$ 10%, 3 phase. 480 VAC $\pm$ 10%, 3 phase.
Line Frequency	50/60 Hz.
Momentary Current	275 Amps at 208 VAC (1 phase). 154 Amps/phase at 208 VAC (3 phase). 250 Amps at 230 VAC (1 phase). 139 Amps/phase at 230 VAC (3 phase). 80 Amps/phase at 400 VAC. 65 Amps/phase at 480 VAC.
Nominal Current *	$\leq$ 5 Amps.
Momentary Power Consumption	55 kVA.

The table below depicts the power requirements for the 50 kW CMP 200<sup>®</sup> / CMP 200<sup>®</sup> DR X-ray generators.

Line Voltage	208 VAC - 5% to 230 VAC + 10%, 3 phase. 400 VAC $\pm$ 10%, 3 phase. 480 VAC $\pm$ 10%, 3 phase.
Line Frequency	50/60 Hz.
Momentary Current	192 Amps/phase at 208 VAC. 174 Amps/phase at 230 VAC. 100 Amps/phase at 400 VAC. 80 Amps/phase at 480 VAC.
Nominal Current *	$\leq$ 5 Amps.
Momentary Power Consumption	65 kVA.

The table below depicts the power requirements for the 65 kW CMP 200<sup>®</sup> DR X-ray generators.

Line Voltage	400 VAC $\pm$ 10%, 3 phase. 480 VAC $\pm$ 10%, 3 phase.
Line Frequency	50/60 Hz.
Momentary Current	125 Amps/phase at 400 VAC. 105 Amps/phase at 480 VAC.
Nominal Current *	$\leq$ 5 Amps.
Momentary Power Consumption	85 kVA.

Use and disclosure is subject to the restrictions on the title page of this CPI document.

### 1.3.2 Generator Power Requirements (Cont)

The table below depicts the power requirements for the 80 kW CMP 200® DR X-ray generators.

Line Voltage	400 VAC $\pm$ 10%, 3 phase. 480 VAC $\pm$ 10%, 3 phase.
Line Frequency	50/60 Hz.
Momentary Current	155 Amps/phase at 400 VAC. 130 Amps/phase at 480 VAC.
Nominal Current *	$\leq$ 5 Amps.
Momentary Power Consumption	105 kVA.

\* Nominal Current = Generator standby current only. External or installer-supplied equipment connected to the generator may increase the nominal current beyond the values shown.

The following table defines the power line requirements for the generators.

**NOTE: THE FOLLOWING TABLE CONTAINS RECOMMENDED VALUES FOR THE WIRE SIZES BETWEEN THE MAINS DISCONNECT AND THE GENERATOR. THE ACTUAL VALUES USED AT AN INSTALLATION ARE DEPENDENT ON THE QUALITY OF THE INPUT LINE (VOLTAGE LEVEL), THE CURRENT REQUIREMENTS, AND THE LENGTH OF THE CABLE RUN, AND MUST BE CONFIRMED BY THE INSTALLER.**

**FINAL SELECTION OF GENERATOR INPUT WIRE AND DISCONNECTS AS WELL AS THE CABLING FROM THE DISTRIBUTION TRANSFORMER TO THE MAINS DISCONNECT MUST MEET THE REQUIREMENTS OF THE LOCAL ELECTRICAL CODES, AND IS USUALLY DETERMINED BY HOSPITAL / CONTRACTOR ENGINEERING.**

**THE RATINGS LISTED CONSIDER THE GENERATOR REQUIREMENTS ONLY. THE INSTALLER MUST MAKE THE NECESSARY COMPENSATION FOR ADDITIONAL LOAD REQUIREMENTS.**

**A POOR QUALITY INPUT LINE MAY RESULT IN THE INSTALLER HAVING TO DERATE THE GENERATOR'S MAXIMUM POWER.**

Generator Series and Mains Voltage	Minimum Recommended Mains Disconnect to Generator (15 ft/5 m max)	Generator Momentary Line Current	Minimum Recommended Generator Service Rating	Minimum Recommended Distribution Transformer Rating	Minimum Recommended Ground Wire Size *	Apparent Mains Resistance
32 kW 208 VAC, 1p.	#2 *** (33 mm <sup>2</sup> )	220 A	120 A	45 kVa	#2 (33 mm <sup>2</sup> )	0.045 $\Omega$
32 kW 230 VAC, 1p.	#2 *** (33 mm <sup>2</sup> )	200 A	120 A	50 kVa	#2 (33 mm <sup>2</sup> )	0.055 $\Omega$
32 kW 208 VAC, 3p.	#4 ** (21 mm <sup>2</sup> )	122 A	100 A	45 kVa	#4 (21 mm <sup>2</sup> )	0.07 $\Omega$
32 kW 230 VAC, 3p.	#4 ** (21 mm <sup>2</sup> )	110 A	100 A	45 kVa	#4 (21 mm <sup>2</sup> )	0.09 $\Omega$
32 kW 400 VAC, 3p.	#6 ** (13.3 mm <sup>2</sup> )	65 A	100 A	45 kVa	#6 (13.3 mm <sup>2</sup> )	0.27 $\Omega$

Use and disclosure is subject to the restrictions on the title page of this CPI document.

## 1.3.2 Generator Power Requirements (Cont)

Generator Series and Mains Voltage	Minimum Recommended Mains Disconnect to Generator (15 ft/5 m max)	Generator Momentary Line Current	Minimum Recommended Generator Service Rating	Minimum Recommended Distribution Transformer Rating	Minimum Recommended Ground Wire Size *	Apparent Mains Resistance
32 kW 480 VAC, 3p.	#6 ** (13.3 mm <sup>2</sup> )	55 A	100 A	45 kVa	#6 (13.3 mm <sup>2</sup> )	0.40 Ω
40 kW 208 VAC, 1p.	#2 *** (33 mm <sup>2</sup> )	275 A	120 A	65 kVa	#2 (33 mm <sup>2</sup> )	0.035 Ω
40 kW 230 VAC, 1p	#2 *** (33 mm <sup>2</sup> )	250 A	120 A	65 kVa	#2 (33 mm <sup>2</sup> )	0.045 Ω
40 kW 208 VAC, 3p.	#4 ** (21 mm <sup>2</sup> )	154 A	100 A	55 kVa	#4 (21 mm <sup>2</sup> )	0.055 Ω
40 kW 230 VAC, 3p.	#4 ** (21 mm <sup>2</sup> )	139 A	100 A	55 kVa	#4 (21 mm <sup>2</sup> )	0.075 Ω
40 kW 400 VAC, 3p.	#6 ** (13.3 mm <sup>2</sup> )	80 A	100 A	55 kVa	#6 (13.3 mm <sup>2</sup> )	0.22 Ω
40 kW 480 VAC, 3p.	#6 ** (13.3 mm <sup>2</sup> )	65 A	100 A	55 kVa	#6 (13.3 mm <sup>2</sup> )	0.32 Ω
50 kW 208 VAC, 3p.	#2 *** (33 mm <sup>2</sup> )	192 A	100 A	65 kVa	#2 (33 mm <sup>2</sup> )	0.045 Ω
50 kW 230 VAC, 3p.	#2 *** (33 mm <sup>2</sup> )	174 A	100 A	65 kVa	#2 (33 mm <sup>2</sup> )	0.055 Ω
50 kW 400 VAC, 3p	#6 ** (13.3 mm <sup>2</sup> )	100 A	100 A	65 kVa	#6 (13.3 mm <sup>2</sup> )	0.17 Ω
50 kW 480 VAC, 3p.	#6 ** (13.3 mm <sup>2</sup> )	80 A	100 A	65 kVa	#6 (13.3 mm <sup>2</sup> )	0.24 Ω
65 kW 400 VAC, 3p	#6 *** (13.3 mm <sup>2</sup> )	125 A	100 A	85 kVa	#6 (13.3 mm <sup>2</sup> )	0.13 Ω
65 kW 480 VAC, 3p	#6 *** (13.3 mm <sup>2</sup> )	105 A	100 A	85 kVa	#6 (13.3 mm <sup>2</sup> )	0.19 Ω
80 kW 400 VAC, 3p	#6 *** (13.3 mm <sup>2</sup> )	155A	100A	105 kVa	#6 (13.3 mm <sup>2</sup> )	0.10 Ω
80 kW 480 VAC, 3p	#6 *** (13.3 mm <sup>2</sup> )	130A	100A	105 kVa	#6 (13.3 mm <sup>2</sup> )	0.15 Ω

\* Refer to 1.3.3 for general grounding information. Maximum wire gauge is # 2 AWG Cu (33 mm<sup>2</sup>).

\*\* Maximum wire gauge is # 4 AWG Cu (21 mm<sup>2</sup>).

\*\*\* Maximum wire gauge is # 2 AWG Cu (33 mm<sup>2</sup>).

Recommended Service Disconnect (as per the above table):

- All wiring and grounding should comply with the national electrical code or equivalent.
- All wiring must be copper.
- The disconnect switch shall be located within reach of the operator.

Use and disclosure is subject to the restrictions on the title page of this CPI document.

### 1.3.3 Generator Ground Requirements

- A suitable ground must be connected from the disconnect switch to the main ground of the generator, located to the right of the main fuse block, on the sub-panel. The ground wire is typically part of the line cord, and the current capacity of the ground conductor must normally be equal to or greater than that of the line conductors.
- A copper ground cable, #10 AWG (6 mm<sup>2</sup>) or larger should be connected from the X-ray tube housing to the High Voltage Module ground stud (located at the top of the High Voltage Module).
- If a neutral line is provided with the system, under no circumstances is it to be used for ground purposes. The ground must carry fault currents only.

### 1.3.4 X-Ray Stator Drive Cable Requirements

The X-ray stator cable used with the starter assembly has to meet the following requirements:

- The cable must be approved to at least 600 VAC if using a dual-speed starter and at least 300 VAC if using a low speed starter.
- The cable must be shielded and the shield ground must be connected to the generator chassis ground and to the tube housing ground.

**Note:** For Metal Centre Section X-ray tubes, **DO NOT** connect the shield ground to the tube housing ground.

- The maximum cable capacitance (from the inner conductors shorted together to the cable shield) **must be less than 5.1 nF**. For example, a cable, type 8618, made by Belden has a capacitance of 4.1 nF for 26 meters.


**Note:** If a longer cable is necessary with larger cable capacitance over 5.1 nF, consult the manufacturer.

Use and disclosure is subject to the restrictions on the title page of this CPI document.

### 1.3.5 Locating and Mounting the Generator

The main generator cabinet is self-standing and does not need to be supported. However, the installation should meet the following requirements:

- The floor must be flat and level.
- The generator installation area must be clean and free of dirt or debris.
- The installer must supply generator hold-down brackets, if required. Alternately, mounting holes have been provided in the base of the generator. The generator may then be anchored to the floor via these holes.
- Sufficient room must be provided to allow access to the rear and side of the generator for installation and service. See figure 1-1 for recommended clearances.
- A cable trough, conduit, or raceway (1 in; 2.5 mm, diameter) should be provided from the control console to the main cabinet to allow routing of the control cable if required.
- The control console is normally freestanding on a desk or shelf. It may be anchored if necessary.



**Figure 1-1: Generator clearances**

Use and disclosure is subject to the restrictions on the title page of this CPI document.

### 1.3.6 Dimensions, Cable Entrance and Seismic Center Location


Figure 1-2 shows the dimensions of the generator cabinet, the locations of the cable access slots, the AC mains cable entry, and the seismic center location for the CMP 200® / CMP 200® DR X-ray generator. The dimensions and weight of the generator and control console(s) are shown in the table below.

ITEM	LENGTH	WIDTH	HEIGHT	WEIGHT
Main cabinet in shipping pack	30.5 (775)*	21.5 (546)*	38 (965)*	151 (68.6)
Main cabinet unpacked	25.7 (653)*	13.7 (348)*	24.3 (617)*	135 (61)
High Voltage Module		12.0 (305)	10.5 (267)	42 (19)
High Voltage Module		12.0 (305)	14.0 (356)	48 (22)
Membrane console	12.3 (313)**	10.9 (277)**	3.7 (94)**	6 (2.72)
Touchscreen control console with base	15.9 (404) maximum**	14.6 (371)**	11.2 (285)**	18 (8.2)
Touchscreen control console without base	13.1 (333)**	14.6 (371)**	3.5 (89)**	10 (4.5)

\* Refer to Figure 1-2

\*\* Refer to Figure 1-3

The above dimensions are inches (mm); weights are in pounds (kg).




CMP200\_013B.CDR

**Figure 1-2: CMP 200® / CMP 200® DR cable entry locations and seismic center**

Use and disclosure is subject to the restrictions on the title page of this CPI document.

## 1.3.6 Dimensions, Cable Entrance, and Seismic Center Location (Cont)



ALL DIMENSIONS ARE IN INCHES (MM)

CMP\_CONSOLES.CDR

**Figure 1-3: CMP 200® / CMP 200® DR operator console dimensions**

## 1.3.7 Tools and Test Equipment Required

The following is a checklist of recommended tools and test equipment for installation and calibration of the generator.

CHECK ✓	DESCRIPTION
	General hand tools for installation: Wrenches, nut drivers, assortment of screwdrivers, pliers, etc.
	If the generator is to be anchored to the floor, suitable tools (i.e. drill, drill bits, etc) and mounting hardware must be available.
	A supply of connectors for wiring: terminal lugs, caps, splices etc.
	A calibrated DVM that indicates true RMS voltages.
	Dual trace memory oscilloscope with a minimum 20 MHz bandwidth; appropriate leads, probes, etc.
	Device for measuring true kVp. This may be a Dynalyzer equivalent or a non-invasive meter such as the Keithley TRIAD system.
	A calibrated radiation meter with detectors that will allow for R/min and uR type measurements (or uGy and Gy/min).
	A suitable mA / mAs meter.
	A strobe or reed type tachometer to verify that the anode is rotating up to speed.
	A sufficient selection of absorbers to allow AEC calibration if this option is fitted. A suggested selection is Lexan in thickness of 5.0, 10.0, and 15.0 cm, or water in plastic containers of homogenous density in thickness of 5.0, 10.0, and 15.0 cm.
	Vapor proof compound for the HV terminations.

Use and disclosure is subject to the restrictions on the title page of this CPI document.




### 1.3.8 Pre-Installation Checklist

Before starting the generator installation, review the following checklist.

CHECK ✓	DESCRIPTION
	Is there an unloading area to transport the generator from the delivery truck to the inside of the building?
	If the installation is not on the same floor as the delivery entrance, is there an elevator available?
	Is there a transport dolly or similar device to move the generator?
	Do any regulatory bodies need to be notified prior to installation?
	If movers are required, have arrangements for time and equipment been completed?
	Are lifting straps or some other suitable device available to lift the generator off the shipping pallet?

### 1.4.0 GENERATOR LAYOUT AND MAJOR COMPONENTS

Figure 1-3 shows the external view and dimensions of the membrane and touchscreen operator's consoles. Figures 1-4 and 1-5 show the major components located inside the generator cabinet. Figure 1-6 is an internal view of the console, showing the major components and cabling. Figure 1-4 does not represent all models. This is meant to show major component layout only.




CMP200\_026B

**Figure 1-4: Major generator subassemblies view 1**

Use and disclosure is subject to the restrictions on the title page of this CPI document.

## 1.4.0 GENERATOR LAYOUT AND MAJOR COMPONENTS (Cont)




CMP200\_007B.CDR


**Figure 1-5: Major generator subassemblies view 2**

Use and disclosure is subject to the restrictions on the title page of this CPI document.

## 1.4.0 GENERATOR LAYOUT AND MAJOR COMPONENTS (Cont)



CMP200\_008B.CDR

**Figure 1-6: Console internal view****Figure 1-7**

Use and disclosure is subject to the restrictions on the title page of this CPI document.

**1.5.0 COMPATIBILITY LISTING**

This X-ray generator is compatible with the following equipment:

**X-RAY TUBES:**

Refer to the compatibility statement and to chapter 2 of this manual.

**NOTE: REFER TO THE SECTIONS “LOW SPEED STARTER TUBE COMPATIBILITY” OR “PROGRAMMING THE DUAL-SPEED STARTER” IN CHAPTER 2 OF THIS MANUAL FOR FURTHER DETAILS.**

**AEC DEVICES:**

Refer to the compatibility statement at the front of this manual.

**DR INTERFACE:**

Refer to the compatibility statement at the front of this manual.

Use and disclosure is subject to the restrictions on the title page of this CPI document.

## 1.6.0 RUNTIME LICENSE AGREEMENT (TOUCHSCREEN CONSOLE)

### User / End User License Agreement

You should carefully read the following terms and conditions before using this product. It contains software ("Software"), the use of which is licensed by Communications & Power Industries Canada Inc. ("CPI") to you, the original end user, for your use only as set forth below. If you do not agree to the terms and conditions of the agreement, do not use the software. If you use any part of the software, such use shall indicate that you accept these terms.

1. The Software embedded in this Product is protected by international intellectual property rights and treaties. You may use the software in object code form only and are prohibited from distributing the software as such. There is no grant of any of the Software's proprietary source code. This license does not grant you any rights to patents, copyright, trade secrets, trademarks, or any other rights with respect to the Software. You are not authorized to modify or to create derivative works based on this software. You are not authorized to decompile, disassemble or otherwise reverse engineer the embedded Software. CPI reserves all rights not expressly granted herein.
2. You are authorized to copy the Software only for the following purposes.
  - a. If such copying is an essential step in the use of the Software and in accordance with this agreement; or
  - b. To make an archival copy for back up purposes.
3. If making a copy pursuant to Section 2 above, you must ensure that all trademark, copyright and intellectual property notices are reproduced and included on any copies that are made.
4. The Software embedded in this Product is provided to you within the same warranty terms, as those offered for the Products, except for any third party software, which is offered "as is" and without warranties of any kind including, but not limited to: warranties of merchantability, fitness for a particular purpose, title and non-infringement.
5. The Software embedded in this Product is not intended for use in or with systems, devices, or products intended to support or sustain life or for any aviation or nuclear reactor application in which the software or its failure, malfunction, or inadequacy could directly or indirectly cause or contribute to personal injury or death or significant property damage. It is your responsibility, as the User or End User, to ascertain the suitability of the Software for your particular situation and / or application.
6. This agreement is effective until terminated. The agreement shall terminate immediately if you fail to adhere to the terms and conditions set forth herein. Upon termination, you must immediately cease all use of the Software and destroy any and all copies of the embedded Software in your possession.
7. This agreement represents the entire agreement between you and CPI with respect to the Software, and supersedes all other agreements or representations, whether written or oral. The terms of this agreement can only be modified by the express written consent of both parties. If any part of this agreement is held to be unenforceable as written, it will be enforced to the maximum extent allowed by applicable law, and without effect to the enforceability of any other part.
8. Should you have questions regarding the use of the Software, please contact [support@cmp.cpii.com](mailto:support@cmp.cpii.com).

Use and disclosure is subject to the restrictions on the title page of this CPI document.

**1.7.0 CUSTOMER SUPPORT**

Address any questions regarding X-ray generator operation to:

Mail: Customer Support Department  
Communications and Power Industries Canada Inc.  
45 River Drive  
Georgetown, Ontario, Canada L7G 2J4

Telephone: (905) 877-0161

Fax: (905) 877-8320  
Attention: Customer Support Department

E-mail: [CANMarketing@.cpil.com](mailto:CANMarketing@.cpil.com)  
Attention: Customer Support Department

**1.8.0 COMPATIBILITY STATEMENT**

The compatibility statement for this generator is located at the front of this manual.

Use and disclosure is subject to the restrictions on the title page of this CPI document.

### 1.9.0 OPEN-SOURCE LICENSE AGREEMENT (TOUCHSCREEN CONSOLE)

The software included in the CPI product covered by this manual may contain copyrighted software programs that are licensed under the General Public License (GPL) V2. A copy of the GPL can be found on the GNU website, [www.gnu.org](http://www.gnu.org). You may obtain the corresponding source code for a period of three years by forwarding your request to:

CPI Communications & Medical Products Division  
45 River Drive  
Georgetown, Ontario, Canada  
L7G 2J4

Telephone: (905) 877-0161

Fax: (905) 877-5327

E-mail: [CANMarketing@cpil.com](mailto:CANMarketing@cpil.com) Subject: Open Source Code

Note: The subject line for any requests must indicate "Open Source Code" in order to obtain a response. Distribution charges applicable at the time of request may apply.

All Open Source Software will be provided "AS IS"; there are (i) no representations or warranties and (ii) neither CPI Canada, nor any of the developers or contributors to Open Source Software shall have any liability or obligation to the customer with respect to Open Source Software beyond what is granted in the particular Open Source Software license. Any modification to software code residing in CPI Canada products shall void all warranties and render products "Not for Clinical Use". CPI Canada shall have no liability or obligation for any CPI Canada products containing modified software.

Address any questions regarding X-ray generator operation to:

Mail: Customer Support Department  
Communications and Power Industries Canada Inc.  
45 River Drive  
Georgetown, Ontario, Canada L7G 2J4

Telephone: (905) 877-0161

Fax: (905) 877-8320  
Attention: Customer Support Department

E-mail: [CANMarketing@cpil.com](mailto:CANMarketing@cpil.com)  
Attention: Customer Support Department

Use and disclosure is subject to the restrictions on the title page of this CPI document.

(This page intentionally left blank)

Use and disclosure is subject to the restrictions on the title page of this CPI document.